

# Efficacy of Bidirectional Fiber-optic Phototherapy for Neonatal Hyperbilirubinemia

K. L. Tan, FRCPE, FRACP

**ABSTRACT.** *Objective.* To evaluate the efficacy of fiber-optic phototherapy using the standard Ohmeda Biliblanket, a large version, double standard Biliblankets, and conventional phototherapy using daylight fluorescent lamps in full-term, healthy infants with nonhemolytic hyperbilirubinemia.

*Methods.* Full-term, healthy infants with nonhemolytic hyperbilirubinemia (bilirubin concentration,  $>255 \mu\text{mol/L}$  or  $222 \mu\text{mol/L}$  at  $<48$  hours of age) were allocated randomly to one of four modes of phototherapy: standard fiber-optic mat (Ohmeda Biliblanket), a large version, double standard Biliblankets, and conventional phototherapy. Bilirubin levels were monitored every 12 hours. Exposure was stopped when bilirubin levels were less than  $185 \mu\text{mol/L}$ , the minimum duration being 24 hours.

*Results.* A total of 171 infants were studied; 42 were exposed to standard fiber-optic phototherapy, 43 to large fiber-optic phototherapy, 42 to double-fiber-optic phototherapy, and 44 to conventional phototherapy. Durations of exposure were  $87.05 \pm 6.09$  (SEM),  $82.57 \pm 5.84$ ,  $64.85 \pm 5.43$ , and  $62.61 \pm 3.74$  hours, respectively; the 24-hour decline rates were  $10.26\% \pm 1.84\%$ ,  $14.50\% \pm 1.53\%$ ,  $21.82\% \pm 1.71\%$ , and  $19.00\% \pm 1.65\%$ , respectively; the overall decline rates over the whole exposure period were  $0.47\% \pm 0.03\%$ ,  $0.52\% \pm 0.04\%$ ,  $0.71\% \pm 0.05\%$ , and  $0.75\% \pm 0.04\%$  per hour, respectively. The efficacy of double-fiber-optic phototherapy and conventional phototherapy was similar and significantly better than that of the large fiber-optic mat and the standard fiber-optic mat in duration, 24-hour decline rate, and overall decline rate. The large mat was slightly better than the standard-size mat with regard to 24-hour decline rate and overall decline rate, but this difference was not significant. Failure of phototherapy occurred only in the large fiber-optic mat group (3 of 43) and the standard fiber-optic mat group (4 of 42); none occurred in the other two groups, but differences not statistically significant. The nursing personnel were more comfortable with single fiber-optic phototherapy, which caused no initial disturbance to the swaddled infants as did conventional phototherapy, but found double-fiber-optic phototherapy difficult to use.

*Conclusion.* For efficacy of fiber-optic phototherapy in full-term infants to be comparable to that of our conventional phototherapy, the light dose of the standard mats needs to be doubled. *Pediatrics* 1997;99(5). URL: <http://www.pediatrics.org/cgi/content/full/99/5/e13>; fiber-optic phototherapy, neonatal hyperbilirubinemia, efficacy.

Fiber-optic phototherapy delivered via a fiber-optic cable to a transparent flat device (mat) that can be placed directly in contact with the infant skin has been demonstrated to be effective for neonatal hyperbilirubinemia.<sup>1-3</sup> The fiber-optic Biliblanket device (Ohmeda Critical Care, Columbia, MD) was found to be more effective than the Wallaby phototherapy system (Fiberoptic Medical Products Inc, Allentown, PA).<sup>4</sup> However, in our experience the efficacy of the standard-size mat for full-term infants is distinctly less than that of conventional phototherapy using our own setup<sup>5</sup>; this was attributed to the relatively small size of the mat, resulting in exposure of the skin being limited to a small area. A bigger mat or two standard mats would improve the performance of fiber-optic phototherapy; the present report compares these two forms of phototherapy against that of a standard fiber-optic mat and conventional phototherapy in terms of efficacy and practicality.

## METHODS

Full-term, healthy infants with nonhemolytic hyperbilirubinemia as previously defined<sup>5</sup> (no abnormality on a hemogram, no evidence of blood group isoimmunization, a negative result of the direct Coombs test, hemoglobin level greater than  $140 \text{ g/L}$ , and hematocrit level greater than 0.40 with exclusion of glucose-6-phosphate dehydrogenase deficiency [tested by a modification of the method of Bernstein by Tan and Boey<sup>6</sup>]) were exposed to phototherapy when their bilirubin concentrations were greater than  $255 \mu\text{mol/L}$  ( $15 \text{ mg/dL}$ ) or greater than  $222 \mu\text{mol/L}$  ( $13 \text{ mg/dL}$ ) in the first 48 hours of life (early onset jaundice). Bilirubin concentrations greater than  $255 \mu\text{mol/L}$  cause significant prolongation of the central conduction time.<sup>7</sup> The infants were randomly allocated using the lottery method to four forms of phototherapy: (1) a standard-size fiber-optic Biliblanket, a device consisting of a halogen lamp with an attached fiber-optic cable containing 2400 optic fibers that end spread out in a flat mat; the light is transmitted via the fibers to the mat, which is placed in direct contact with the skin during phototherapy; (2) a single large fiber-optic mat, a stretched version of the standard mat with the same number of optic fibers; (3) double fiber-optic mats, two of the standard mats, one placed against the front and the other against the back of the infant; and (4) conventional phototherapy using seven overhead daylight fluorescent lamps (TLD18W/54; Philips Electronic Instruments, Mahwah, NJ) arranged in an arc 35 cm above the infant, a height that permitted clear observation of, as well as good accessibility to, the infants and at the same time provided adequate heat to maintain normothermia. In the first three groups it was possible to swaddle the infants with the mat(s) placed against the infants' skin; to ensure maximal efficacy, the fiber-optic mat was used without its sheath and set at maximal power. The size of the mat was  $11 \times 20 \text{ cm}$ , and the illuminated part was  $11 \times 13 \text{ cm}$ ; those of the large mat were  $11 \times 24$  and  $11 \times 16 \text{ cm}$ , respectively. The increment in size of the illuminated part was about 23%. No eye pads were required. In the conventional phototherapy group, the infants were exposed completely unclothed, with their eyes covered.

The irradiance of the standard fiber-optic device (without the

From the Department of Neonatology, National University Hospital, Singapore.

Received for publication Sep 5, 1996; accepted Sep 5, 1996.

Reprint requests to (K.L.T.) Department of Neonatology, National University Hospital, 5 Lower Kent Ridge Rd, Singapore 119074.

PEDIATRICS (ISSN 0031 4005). Copyright © 1997 by the American Academy of Pediatrics.

sheath and set at maximal power) measured at the center and the four corners averaged 19.01  $\mu\text{W}/\text{cm}^2$  per nanometer; that of the seven overhead lamps averaged 6.73  $\mu\text{W}/\text{cm}^2$  per nanometer; the irradiance of the larger fiber-optic mat was similar to that of the standard mat; because of its larger size, the light dose of the larger mat would be about 23% more than that of the standard fiber-optic mat. Because of the different spectra of the two types of light, the irradiance values of the Biliblanket were 867  $\mu\text{W}/\text{cm}^2$  in the 400- to 480-nm range, 437.0  $\mu\text{W}/\text{cm}^2$  in the 425- to 475-nm range, 342.0  $\mu\text{W}/\text{cm}^2$  in the 440- to 480-nm range, and 775.8  $\mu\text{W}/\text{cm}^2$  in the 440- to 500-nm range; the values of the seven overhead lamps were 403.2, 205, 106.6, and 201.6  $\mu\text{W}/\text{cm}^2$ , respectively. In the double-fiber-optic setup the total irradiance would be twice that of the standard fiber-optic mat. The irradiance in the blue spectrum was relatively low, but that in green spectrum was substantial in fiber-optic phototherapy (hence the greater values in the spectrum involving the 500-nm band). The measurements were made using an International Light (Newburyport, MA) 400A radiometer/photometer. Fluid intake was increased during phototherapy to offset the increased fluid loss during exposure.

Capillary blood was sampled at start of exposure and every 12 hours thereafter to monitor the serum bilirubin response to exposure. The lights were switched off during sampling. The capillary samples were placed in labeled red drinking straws and kept in a light-proof box until the moment of determination under standard conditions using an American Optical bilirubinometer, which was calibrated regularly against known standards.

In infants with increasing bilirubin values exceeding the starting value on two consecutive determinations during exposure, direct-acting bilirubin was determined as previously described;<sup>6</sup> when this was minimal (<10  $\mu\text{mol}/\text{L}$  [0.6 mg/dL]), phototherapy was deemed to have failed, and the infant was transferred to high-intensity phototherapy as previously described.<sup>8</sup> Direct-acting bilirubin was also estimated in random samples.

Phototherapy was terminated when bilirubin values had declined to less than 185  $\mu\text{mol}/\text{L}$  (11 mg/dL) on two successive estimations, the minimal duration being 24 hours; the prolonged central conduction time was observed to improve with bilirubin decline during exposure, complete reversal occurring 24 hours after cessation of exposure.<sup>7</sup> Bilirubin levels were then monitored daily to determine the rebound, for at least 2 days; if rebound bilirubin concentrations increased beyond those of the prephototherapy values, additional phototherapy was performed following the same guidelines.

The nursing personnel caring for the infants were interviewed regarding convenience, ease of use, infant care, and acceptability of the fiber-optic device compared with conventional phototherapy. The parents were also interviewed regarding their reaction to the new device. Informed consent from the parent(s) and approval for this study from the director of medical affairs of the National University Hospital were obtained.

The data were statistically evaluated using analysis of variance, Student's *t* test, and the  $\chi^2$  test.

## RESULTS

Altogether 171 full-term, healthy infants (Table 1) with nonhemolytic hyperbilirubinemia were stud-

ied. All remained well during and after the exposure; phototherapy was well tolerated, but initial restlessness was observed with conventional phototherapy before the infants settled down. Phototherapy was effective in decreasing bilirubin concentrations in all four groups. The response was the greatest in the double-fiber-optic and daylight groups, the former demonstrating slightly better efficacy initially, with the latter being better in overall efficacy (Fig 1). Statistical evaluation by analysis of variance demonstrated significant differences among the four groups in duration of exposure ( $P < .003$ ), 24-hour decline ( $P < .001$ ), and decline over the whole exposure duration ( $P < .001$ ). Student's *t* test was then performed to evaluate the differences between individual groups. The large fiber-optic mat was slightly better than the standard-size fiber-optic mat, but the improvement was not significant; duration of exposure was about 5% shorter, 24-hour decline rate (expressed as percentage of starting bilirubin concentration) 40% better, and overall decline rate (decline during the whole period of exposure expressed as percentage of decline per hour) 10% faster. Both of these phototherapy modes were significantly less effective than double-fiber-optic and daylight phototherapy (Table 2). Failure occurred only with the standard and large fiber-optic mats; none occurred with the double-fiber-optic and conventional modes, but the difference was not significant. One infant each from the double phototherapy and conventional phototherapy groups needed second exposures compared with none in the other two groups. The response to the second exposure was as good as to the first.

The direct-acting bilirubin was minimal in all the samples tested; all the values obtained were less than 10  $\mu\text{mol}/\text{L}$ .

The nurses caring for the full-term infants were unanimous in their approval of the fiber-optic mat, being more comfortable without the conventional overhead phototherapy frame; absence of glare was a positive factor, although this was thought to be a minor problem. Cleaning the soiled fiber-optic mats was the only disadvantage, but this was offset by little need to clean the cots, otherwise required during conventional phototherapy. The nurses thought

TABLE 1. Data of Infants Studied

Infant Characteristic	Daylight Phototherapy	Fiber-optic Phototherapy		
		Double Mat	Large Mat	Standard Mat
No. (M:F)	44 (27:17)	42 (20:22)	43 (26:17)	42 (23:19)
Birth weight, g*	3080.5 $\pm$ 77.1	3119.2 $\pm$ 64.5	3059.4 $\pm$ 57.7	3200.1 $\pm$ 49.8
Gestational age, wk*	38.2 $\pm$ 0.3	38.7 $\pm$ 0.2	38.8 $\pm$ 0.2	38.6 $\pm$ 0.2
Age, d*	4.1 $\pm$ 0.2	4.0 $\pm$ 0.2	4.1 $\pm$ 0.2	4.0 $\pm$ 0.2
Hemoglobin, g/L*				
Start	181.5 $\pm$ 3.4	184.2 $\pm$ 3.0	180.5 $\pm$ 2.9	185.0 $\pm$ 4.1
End	169.4 $\pm$ 3.1	164.2 $\pm$ 5.0	165.9 $\pm$ 2.8	167.3 $\pm$ 3.1
Hematocrit*				
Start	0.56 $\pm$ 0.01	0.56 $\pm$ 0.01	0.57 $\pm$ 0.03	0.56 $\pm$ 0.01
End	0.51 $\pm$ 0.01	0.51 $\pm$ 0.01	0.51 $\pm$ 0.01	0.51 $\pm$ 0.01
Bilirubin, $\mu\text{mol}/\text{L}$ *				
Start	261.8 $\pm$ 2.7	259.8 $\pm$ 2.9	262.8 $\pm$ 2.5	260.9 $\pm$ 2.4
End	152.8 $\pm$ 2.6	153.8 $\pm$ 2.6	164.0 $\pm$ 2.2	165.0 $\pm$ 2.2

\* Values expressed as mean  $\pm$  SEM.



bank blue-light phototherapy<sup>13</sup> used at an optimal dose<sup>11,13</sup> should provide maximal efficacy in full-term infants with severe or rapidly increasing jaundice,<sup>15</sup> especially of a hemolytic nature.

This study demonstrated that fiber-optic phototherapy with the Ohmeda Biliblanket in full-term infants was adequate for routine use only with the double arrangement, sandwiching the infant in between two mats. This apparent inadequacy of the standard fiber-optic Biliblanket for full-term infants reinforces our earlier study,<sup>5</sup> in which exposure was too long, and failures occurred too often. Increasing the light dose and improving the spectral emission of the lamp used will enhance efficacy. If spectral emission were not improved, an increase in the light dose of 100% would be needed for performance comparable to that of conventional phototherapy; this would require an increase of light intensity, blanket size, or both. When the spectral emission of the light can also be improved, the overall light dose need not be increased to this degree to ensure the same result; a single enlarged mat with a more appropriate light spectrum might thus be comparable to conventional phototherapy in efficacy. When such a situation can be realized, then fiber-optic phototherapy, with its advantages of convenience, ease of use, freedom from obstruction, and easy accessibility, might then be able to achieve its full potential and be the preferred choice for treating neonatal hyperbilirubinemia. Until then, the present Ohmeda Biliblanket should mainly be used for small preterm infants who generally would respond adequately to such treatment.<sup>5</sup>

#### ACKNOWLEDGMENTS

I thank the chief executive officer and director of medical affairs for permission to conduct this study, the medical and nursing

personnel for their enthusiastic support, L. E. Lee and P. Y. Chua for technical support, and L. H. Ang for secretarial assistance.

#### REFERENCES

1. Rosenfeld W, Twist P, Concepcion L. A new device for phototherapy treatment of jaundiced infants. *J Perinatol.* 1990;10:243-247
2. Holtrop PC, Madison K, Maisels MJ. A clinical trial of fiberoptic phototherapy vs conventional phototherapy. *Am J Dis Child.* 1992;146:235-237
3. Holtrop PC, Ruedisueli K, Maisels JF. Double versus single phototherapy in low birth weight infants. *Pediatrics.* 1992;90:674-677
4. George P, Lynch M. Ohmeda Biliblanket vs Wallaby phototherapy system for the reduction of bilirubin levels in the home care setting. *Clin Pediatr.* 1994;33:178-180
5. Tan KL. Comparison of the efficacy of fiberoptic and conventional phototherapy for neonatal hyperbilirubinemia. *J Pediatr.* 1994;125:607-612
6. Tan KL, Boey KW. Efficacy of phototherapy on neonatal hyperbilirubinaemia. *Br Med J.* 1986;293:1361-1364
7. Tan KL, Skurr BA, Yip YY. Phototherapy and the brain-stem auditory evoked response in neonatal hyperbilirubinemia. *J Pediatr.* 1992;120:306-308
8. Tan KL. The pattern of bilirubin response to phototherapy for neonatal hyperbilirubinemia. *Pediatr Res.* 1982;16:670-674
9. Tan KL. Efficacy of fluorescent daylight, blue and green lamps in the management of non-hemolytic hyperbilirubinemia. *J Pediatr.* 1989;114:132-137
10. Donzelli GP, Pratesi R, Rapisardi G, et al. 1-day phototherapy of neonatal jaundice with blue-green lamp. *Lancet.* 1996;346:184-185
11. Tan KL, Lim GC, Boey KW. Efficacy of high intensity blue light and standard daylight phototherapy for non-haemolytic hyperbilirubinaemia. *Acta Paediatr.* 1992;81:870-874
12. Tan KL. Comparison of the effectiveness of single-direction and double-direction phototherapy for neonatal jaundice. *Pediatrics.* 1975;56:550-553
13. Tan KL. Phototherapy for neonatal jaundice. *Clin Perinatol.* 1991;18:423-429
14. Vogl TP. Phototherapy of neonatal hyperbilirubinemia: bilirubin in unexposed areas of the skin. *J Pediatr.* 1974;85:707-711
15. Tan KL, Lim GC, Boey KW. Phototherapy for ABO haemolytic hyperbilirubinaemia. *Biol Neonate.* 1992;61:358-365

## Efficacy of Bidirectional Fiber-optic Phototherapy for Neonatal Hyperbilirubinemia

K. L. Tan

*Pediatrics* 1997;99:e13

DOI: 10.1542/peds.99.5.e13

### Updated Information & Services

including high resolution figures, can be found at:  
<http://pediatrics.aappublications.org/content/99/5/e13>

### References

This article cites 15 articles, 3 of which you can access for free at:  
<http://pediatrics.aappublications.org/content/99/5/e13#BIBL>

### Permissions & Licensing

Information about reproducing this article in parts (figures, tables) or in its entirety can be found online at:  
<http://www.aappublications.org/site/misc/Permissions.xhtml>

### Reprints

Information about ordering reprints can be found online:  
<http://www.aappublications.org/site/misc/reprints.xhtml>

# American Academy of Pediatrics

DEDICATED TO THE HEALTH OF ALL CHILDREN®





# PEDIATRICS®

OFFICIAL JOURNAL OF THE AMERICAN ACADEMY OF PEDIATRICS

## **Efficacy of Bidirectional Fiber-optic Phototherapy for Neonatal Hyperbilirubinemia**

K. L. Tan

*Pediatrics* 1997;99:e13

DOI: 10.1542/peds.99.5.e13

The online version of this article, along with updated information and services, is located on the World Wide Web at:

<http://pediatrics.aappublications.org/content/99/5/e13>

Pediatrics is the official journal of the American Academy of Pediatrics. A monthly publication, it has been published continuously since 1948. Pediatrics is owned, published, and trademarked by the American Academy of Pediatrics, 345 Park Avenue, Itasca, Illinois, 60143. Copyright © 1997 by the American Academy of Pediatrics. All rights reserved. Print ISSN: 1073-0397.

American Academy of Pediatrics

DEDICATED TO THE HEALTH OF ALL CHILDREN®

